

Sultanate of Oman

Ministry of Health

Directorate General of Pharmaceutical Affairs
and Drug Control

MUSCAT



سلطنة عمان
وزارة الصحة
المديرية العامة للصيدل
والرقابة الدوائية
مسقط

To:

THE DIRECTOR GENERAL OF HEALTH SERVICES IN ALL GOVERNORATES
Commanding Officer, Armed Forces Hospital (Al Khoudh & Salalah)
Director General of Engineering Affairs, MOH
Director General of Royal Hospital
Director General of Khoula Hospital
Director General of Medical Supplies (MOH)
Director General of Pvt. Health Est. Affairs (to kindly arrange distribution to all Pvt. Hospitals)
Hospital Director (Al Nahda Hospital)
Hospital Director (Al Massara Hospital)
The Head of Medical Services in SQU Hospital
The Head of Medical Services in Royal Oman Police
The Head of Medical Services in Ministry of Defence
The Head of Medical Services in The Diwan
The Head of Medical Services in The Sultan's Special Force
The Head of Medical Services in Internal Security Services
The Head of Medical Services in Petroleum Development of Oman
The Head of Medical Services in LNG Oman
ALL PRIVATE PHARMACIES & DRUG STORES

After Compliments,

Please find attached our Circular No. ^{12/21}..... dated ^{18/1/21}..... Regarding NCMDR Recall of Face Mask from (Sondos Al debajah Co).

Copy to:


- Director, Office of H.E. The Undersecretary for Health Affairs
- Director of Medical Device Control, DGPA&DC
- Director of Pharmacovigilance & Drug Information Dept, DGPA&DC
- Director of Drug Control Department, DGPA&DC
- Director of Pharmaceutical Licensing Department, DGPA&DC
- Director of Central Quality Control Lab., DGPA&DC
- Supdt. of Central Drug Information

Sultanate of Oman

Ministry of Health

Directorate General of Pharmaceutical Affairs
and Drug Control

MUSCAT



سلطنة عمان
وزارة الصحة
المديرية العامة للأدوية
والرقابة الدوائية
مسقط

Circular No. 12 / 2021

04-06 -1442 H

18 -01-2021

Recall of Face Mask from Sondos Al debajah Co.

Source	NCMDR- National Centre Medical Device Reporting https://ncmdr.sfda.gov.sa/Secure/CA/CaViewRecall.aspx?caid=4&rid=15504
Product	Face Mask.
Manufacturer	Sondos Al debajah Co, KSA.
The affected products	Product Code: SDSF0150T Lot No. 402020
Reason	Many masks found contaminated with black ash. Nose plastic piece is not sealed well which may release from the mask.
Action	1. Please Identify the affected products and stop use it. Manufacturer will contact all the customers and replace all affected products. 2. Contact the local agent for remedial action.
comments	Healthcare professionals are encouraged to report any adverse events Suspected to be associated with the above device or any other medical device to Department of Medical Device Control contact E-mail: Med-device@moh.gov.om

Dr. Mohammed Hamdan Al Rubaie

DIRECTOR GENERAL

