

To:

THE DIRECTOR GENERAL OF HEALTH SERVICES IN ALL GOVERNORATES
Commanding Officer, Armed Forces Hospital (Al Khoudh & Salalah)
Director General of Engineering Affairs, MOH
Director General of Royal Hospital
Director General of Khoula Hospital
Director General of Medical Supplies (MOH)
Director General of Pvt. Health Est. Affairs (to kindly arrange distribution to all Pvt. Hospitals)
Hospital Director (Al Nahda Hospital)
Hospital Director (Al Massara Hospital)
The Head of Medical Services in SQU Hospital
The Head of Medical Services in Royal Oman Police
The Head of Medical Services in Ministry of Defence
The Head of Medical Services in The Diwan
The Head of Medical Services in The Sultan's Special Force
The Head of Medical Services in Internal Security Services
The Head of Medical Services in Petroleum Development of Oman
The Head of Medical Services in LNG Oman
ALL PRIVATE PHARMACIES & DRUG STORES


After Compliments,

Please find attached our Circular No. ^{190/21} dated ^{1/11/2021} Regarding NCMDR Field Safety Corrective Action of Sensis, Sensis Vibe Combo, Sensis Vibe Hemo from (mfr: Siemens).

Copy to:

- Director, Office of H.E. The Undersecretary for Health Affairs
- Director of Medical Device Control, DGPA&DC
- Director of Pharmacovigilance & Drug Information Dept, DGPA&DC
- Director of Drug Control Department, DGPA&DC
- Director of Pharmaceutical Licensing Department, DGPA&DC
- Director of Central Quality Control Lab., DGPA&DC
- Supdt. of Central Drug Information





Circular No. 190/2021

25-03-1443 H

01-11-2021

بتقدم بثقة
Moving Forward
with Confidence



Field Safety Corrective Action of Sensis, Sensis Vibe Combo, Sensis Vibe Hemo from Siemens.

Source	NCMDR- National Centre for Medical Devices Reporting- SFDA. https://ncmdr.sfda.gov.sa/Secure/CA/CAViewRecall.aspx?caid=6&rid=15875
Product	Sensis, Sensis Vibe Combo, Sensis Vibe Hemo.
Description	Radiological technology - radiological equipment for vascular diagnostics.
Manufacturer	Siemens.
Local agent	Muscat Pharmacy & Stores LLC.
The affected products	system currently on VD1X software.
Reason	Missing regular reboot of the system may result in frozen vital signs.
Action	1. The software update will mitigate the occurrence of the issue. 2. Contact the local agent for remedial action.
comments	Healthcare professionals are encouraged to report any adverse events Suspected to be associated with the above device or any other medical device to Department of Medical Device Control through the E-mail: Med-device@moh.gov.om

Dr. Mohammed Hamdan Al Rubaie
Director General

