

# Sultanate of Oman

Ministry of Health

Directorate General of Pharmaceutical Affairs  
and Drug Control

MUSCAT



سلطنة عمان  
وزارة الصحة  
الديرة العامة للصيدل  
والرقابة الدوائية  
مسقط

To:

**THE DIRECTOR GENERAL OF HEALTH SERVICES IN ALL GOVERNORATES**

Commanding Officer, Armed Forces Hospital (Al Khoudh & Salalah)

Director General of Engineering Affairs, MOH

Director General of Royal Hospital

Director General of Khoula Hospital

Director General of Medical Supplies (MOH)

Director General of Pvt. Health Est. Affairs (to kindly arrange distribution to all Pvt. Hospitals)

Hospital Director (Al Nahda Hospital)

Hospital Director (Al Massara Hospital)

The Head of Medical Services in SQU Hospital

The Head of Medical Services in Royal Oman Police

The Head of Medical Services in Ministry of Defence

The Head of Medical Services in The Diwan

The Head of Medical Services in The Sultan's Special Force

The Head of Medical Services in Internal Security Services

The Head of Medical Services in Petroleum Development of Oman

The Head of Medical Services in LNG Oman

**ALL PRIVATE PHARMACIES & DRUG STORES**

After Compliments,

Please find attached our Circular No. 195 dated 21/10/20 Regarding NCMDR Recall of Langston Dual Lumen Catheter from (mfr: Teleflex Medical).

Copy to:

- Director, Office of H.E. The Undersecretary for Health Affairs
- Director of Medical Device Control, DGPA&DC
- Director of Pharmacovigilance & Drug Information Dept, DGPA&DC
- Director of Drug Control Department, DGPA&DC
- Director of Pharmaceutical Licensing Department, DGPA&DC
- Director of Central Quality Control Lab., DGPA&DC
- Supdt. of Central Drug Information

# Sultanate of Oman

Ministry of Health

Directorate General of Pharmaceutical Affairs  
and Drug Control

MUSCAT



سلطنة عمان  
وزارة الصحة  
المديرية العامة للأدوية  
والسراويل والصيدلانية  
مسقط

Circular No. 195/2020

04 -02-1442 H

21 -10-2020

## Recall of Langston Dual Lumen Catheter from Teleflex Medical.

Source	National Centre for Medical Devices Reporting- NCMDR <a href="https://ncmdr.sfda.gov.sa/Secure/CA/CaViewRecall.aspx?caid=4&amp;rid=15380">https://ncmdr.sfda.gov.sa/Secure/CA/CaViewRecall.aspx?caid=4&amp;rid=15380</a>
Product	Langston Dual Lumen Catheter.
Description	Injections / Infusions / Transfusions / Dialysis – catheters.
Manufacturer	Teleflex Medical.
Local agent	Muscat Pharmacy & Stores LLC.
The affected products	Model No. 5515, 5545, 5550, and 5540 Multiple lot numbers of the affected device are provided in the attached FSN.
Reason	Inner Lumen May Separate from Device Hub during or after a Powered Contrast Injection.
Action	1. Discontinue use of affected catheters. 2. Contact the local agent for remedial action.
Product image	
comments	Healthcare professionals are encouraged to report any adverse events Suspected to be associated with the above device or any other medical device to Department of Medical Device Control through an E-mail: <a href="mailto:Med-device@moh.gov.om">Med-device@moh.gov.om</a>

Dr. Mohammed Hamdan Al Rubaie

Director General

