

Sultanate of Oman

Ministry of Health

Directorate General of Pharmaceutical Affairs
and Drug Control

MUSCAT



سِلاطِنَا اَمَانَا
وَزَارَةِ الصِّحَّةِ
وَالرِّقَابَةِ الدَّوَلِيَّةِ
مَسْقَط

To:

THE DIRECTOR GENERAL OF HEALTH SERVICES IN ALL GOVERNORATES

Commanding Officer, Armed Forces Hospital (Al Khoudh & Salalah)

Director General of Engineering Affairs, MOH

Director General of Royal Hospital

Director General of Khoula Hospital

Director General of Medical Supplies (MOH)

Director General of Pvt. Health Est. Affairs (to kindly arrange distribution to all Pvt. Hospitals)

Hospital Director (Al Nahda Hospital)

Hospital Director (Al Massara Hospital)

The Head of Medical Services in SQU Hospital

The Head of Medical Services in Royal Oman Police

The Head of Medical Services in Ministry of Defence

The Head of Medical Services in The Diwan

The Head of Medical Services in The Sultan's Special Force

The Head of Medical Services in Internal Security Services

The Head of Medical Services in Petroleum Development of Oman

The Head of Medical Services in LNG Oman

ALL PRIVATE PHARMACIES & DRUG STORES

After Compliments,

Please find attached our Circular No. 49..... dated 02/03/20 Regarding recall of Arrow Epidural Catheterization Kits and Sets from Teleflex Medical

Copy to:

- Director, Office of H.E. The Undersecretary for Health Affairs
- Director of Medical Device Control, DGPA&DC
- Director of Pharmacovigilance & Drug Information Dept, DGPA&DC
- Director of Drug Control Department, DGPA&DC
- Director of Pharmaceutical Licensing Department, DGPA&DC
- Director of Central Quality Control Lab., DGPA&DC
- Supdt. of Central Drug Information

Sultanate of Oman

Ministry of Health

Directorate General of Pharmaceutical Affairs
and Drug Control

MUSCAT

Circular No. 49 / 2020

07-7-1441 H


02-03-2020



سلطنة عمان
وزارة الصحة
المديرية العامة للأدوية
والرقابة الدوائية
مسقط

Ref: 33/2020

Recall of Arrow Epidural Catheterization Kits and Sets from Teleflex Medical

Source of Recall	NCMDR National Centre for Medical Devices Reporting, SFDA https://ncmdr.sfda.gov.sa/Secure/CA/CAViewRecall.aspx?caid=6&rid=15072
Product	Arrow Epidural Catheterization Kits and Sets
Manufacturer	Teleflex Medical
Local Agent	Muscat Pharmacy
The affected products	AA-05400-B, ASK-05500-TM, BE-05400BETTEL, EC-05520-P, LR-05501, UR-05501
Reason for Recall	LOR Syringes May Fail, Potentially Resulting in Dural Puncture
Action	1. Kindly check your stock, contact your local agent for remedial action
Product image	
comments	Healthcare professionals are encouraged to report any adverse events Suspected to be associated with the above device or any other medical Device to Director of Medical Device Control contact E-mail dg-padc@moh.gov.om

Directorate General of Pharmaceutical Affairs & Drug Control
Sultanate of Oman

Dr. Mohammed Hamdan Al Rubaie
DIRECTOR GENERAL

