



To:

THE DIRECTOR GENERAL OF HEALTH SERVICES IN ALL GOVERNORATES

Commanding Officer, Armed Forces Hospital (Al Khoudh & Salalah)

Director General of Engineering Affairs, MOH

Director General of Royal Hospital

Director General of Khoula Hospital

Director General of Medical Supplies (MOH)

Director General of Pvt. Health Est. Affairs (to kindly arrange distribution to all Pvt. Hospitals)

Hospital Director (Al Nahda Hospital)

Hospital Director (Al Massara Hospital)

The Head of Medical Services in SQU Hospital

The Head of Medical Services in Royal Oman Police

The Head of Medical Services in Ministry of Defence

The Head of Medical Services in The Diwan

The Head of Medical Services in The Sultan's Special Force

The Head of Medical Services in Internal Security Services

The Head of Medical Services in Petroleum Development of Oman

The Head of Medical Services in LNG Oman

ALL PRIVATE PHARMACIES & DRUG STORES

After Compliments,

Please find attached our Circular No...103.... dated 24/5/22 Regarding NCMDR FSN of RAMYPLUS 3 way stopcock from (mrf: Alshifa Medical Products Co CJS).

Copy to:

- Director, Office of H.E. The Undersecretary for Health Affairs
- Director of Medical Device Control, DGPA&DC
- Director of Pharmacovigilance & Drug Information Dept, DGPA&DC
- Director of Drug Control Department, DGPA&DC
- Director of Pharmaceutical Licensing Department, DGPA&DC
- Director of Central Quality Control Lab., DGPA&DC
- Supdt. of Central Drug Information



Circular No. 103 / 2022

ببندفم
Forward
with Confidence



25 -10-1443 H

24 -05-2022

Field Safety Notice of RAMYPLUS 3 way stopcock from Alshifa Medical Products Co CJS

Source	NCMDR-National Center for Medical Device Reporting https://ncmdr.sfda.gov.sa/Secure/CA/CaViewRecall.aspx?caid=4&rid=16131
Product	RAMYPLUS 3 way stopcock.
Description	Stopcock.
Manufacturer	Alshifa Medical Products Co CJS.
Local Agent	Salalah Medical Supplies Manufacturing Co. LLC.
The affected products	3-way stopcock with 10cm tube. Ref: F20317200101
Reason	Tube disconnected leading to blood loss during use.
Action	1. Use the appropriate 3 way stopcock for your application. This 3-way stopcock shouldn't be used in high pressure application. 2. Read and understand the IFU or labels properly. 3. Dispose of the defective device and replace with another. 4. Contact the local agent for remedial action.
comments	Healthcare professionals are encouraged to report any adverse events Suspected to be associated with the above device or any other medical device to Department of Medical Device Control contact E-mail: Med-device@moh.gov.om

Dr. Mohammed Hamdan Al-Rubaie
DIRECTOR GENERAL

