


# Sultanate of Oman

Ministry of Health

Directorate General of Pharmaceutical Affairs  
and Drug Control



سِلاطِنَا اَمَانَا  
وَزَارَةِ الصِّحَّةِ  
وَالرِّقَابَةِ الدَّوَلِيَّةِ  
مَسْقُوطَا

MUSCAT

To:

**THE DIRECTOR GENERAL OF HEALTH SERVICES IN ALL GOVERNORATES**

Commanding Officer, Armed Forces Hospital (Al Khoudh & Salalah)

Director General of Engineering Affairs, MOH

Director General of Royal Hospital

Director General of Khoula Hospital

Director General of Medical Supplies (MOH)

Director General of Pvt. Health Est. Affairs (to kindly arrange distribution to all Pvt. Hospitals)

Hospital Director (Al Nahda Hospital)

Hospital Director (Al Massara Hospital)

The Head of Medical Services in SQU Hospital

The Head of Medical Services in Royal Oman Police

The Head of Medical Services in Ministry of Defence

The Head of Medical Services in The Diwan

The Head of Medical Services in The Sultan's Special Force

The Head of Medical Services in Internal Security Services

The Head of Medical Services in Petroleum Development of Oman

The Head of Medical Services in LNG Oman

**ALL PRIVATE PHARMACIES & DRUG STORES**

After Compliments,

Please find attached our Circular No. 148... dated 20/08/20... Regarding NCMDR Recall of Transport Medium from (mfr: Citotest Labware Manufacturing CO.,LTD).

Copy to:

- Director, Office of H.E. The Undersecretary for Health Affairs
- Director of Medical Device Control, DGPA&DC
- Director of Pharmacovigilance & Drug Information Dept, DGPA&DC
- Director of Drug Control Department, DGPA&DC
- Director of Pharmaceutical Licensing Department, DGPA&DC.
- Director of Central Quality Control Lab., DGPA&DC
- Supdt. of Central Drug Information

# Sultanate of Oman

Ministry of Health


Directorate General of Pharmaceutical Affairs  
and Drug Control

MUSCAT

Circular No. 148/ 2020


30 -12-1441 H

20 -08-2020



سُلْطَانَاةُ عُومَانِ  
وَزَارَاةُ الصِّحَّةِ  
وَالرِّيَاةِ الْعَامَّةِ وَالصِّدْقِ  
وَالرَّقَابَةِ الدَّوْلِيَّةِ  
مَسْقَط

## Recall of Transport Medium from Citotest Labware Manufacturing CO.,LTD.

Source	NCMDR- National Centre Medical Device Reporting <a href="https://ncmdr.sfda.gov.sa/Secure/CA/CAViewRecall.aspx?caid=4&amp;rid=15284">https://ncmdr.sfda.gov.sa/Secure/CA/CAViewRecall.aspx?caid=4&amp;rid=15284</a>
Product	Transport Medium-In vitro diagnostic devices.
Manufacturer	Citotest Labware Manufacturing CO.,LTD
Local Agent	Global Source Trading LLC.
The affected products	Lot No. 042001 Affected serial numbers: 2118-0019
Reason	The virus specimens transportation tube may leak.
Action	1. Affected products will be replaced. 2. Contact the local agent for remedial action.
Product image	
comments	Healthcare professionals are encouraged to report any adverse events Suspected to be associated with the above device or any other medical Device to Department of Medical Device Control contact E-mail: <a href="mailto:Med-device@moh.gov.om">Med-device@moh.gov.om</a>



Dr. Mohammed Hamdan Al Rubaie

DIRECTOR GENERAL