

Sultanate of Oman

Ministry of Health

Directorate General of Pharmaceutical Affairs
and Drug Control
MUSCAT



سلطنة عمان
وزارة الصحة
المديرية العامة للصيدل
والرقابة الدوائية
مسقط

To:

THE DIRECTOR GENERAL OF HEALTH SERVICES IN ALL GOVERNORATES

Commanding Officer, Armed Forces Hospital (Al Khoudh & Salalah)

Director General of Engineering Affairs, MOH

Director General of Royal Hospital

Director General of Khoula Hospital

Director General of Medical Supplies (MOH)

Director General of Pvt. Health Est. Affairs (to kindly arrange distribution to all Pvt. Hospitals)

Hospital Director (Al Nahda Hospital)

Hospital Director (Al Massara Hospital)

The Head of Medical Services in SQU Hospital

The Head of Medical Services in Royal Oman Police

The Head of Medical Services in Ministry of Defence

The Head of Medical Services in The Diwan

The Head of Medical Services in The Sultan's Special Force

The Head of Medical Services in Internal Security Services

The Head of Medical Services in Petroleum Development of Oman

The Head of Medical Services in LNG Oman

ALL PRIVATE PHARMACIES & DRUG STORES

After Compliments,

Please find attached our Circular No...23..... dated 29/01/12 Regarding Health council

Recall of Arrow EZ-IO-Intraosseous vascular access needle sets (Mfr: Teleflex Medical)

Copy to:

- Director, Office of H.E. The Undersecretary for Health Affairs
- Director of Medical Device Control, DGPA&DC
- Director of Pharmacovigilance & Drug Information Dept, DGPA&DC
- Director of Drug Control Department, DGPA&DC
- Director of Pharmaceutical Licensing Department, DGPA&DC
- Director of Central Quality Control Lab., DGPA&DC
- Supdt. of Central Drug Information

Sultanate of Oman

Ministry of Health

Directorate General of Pharmaceutical Affairs
and Drug Control

MUSCAT



سلطنة عمان
وزارة الصحة
الديرة العامة للأدوية
والرقابة الدوائية
مسقط


Circular No. 23 / 2020

Ref: 16/2020

04 -06 1441 H

29 -01-2020

Recall of Arrow EZ-IO from Teleflex Medical

Source of Recall	Gulf Health council file:///C:/Users/moh76697/Downloads/Intraosseous%20vascular%20access%20needle%20sets.pdf
Product	Arrow EZ-IO-Intraosseous vascular access needle sets
Manufacturer	Teleflex Medical
Local agent	Muscat pharmacy & Stores L.L.C
The affected products	9018P-NO-005; 9018P-VC-005; 9001P-EE-005; 9001P-EU-005; 9001P-ME-005; 9001P-NO-005; 9001P-VC-005; 9079P-EE-005; 9079P-EU-005; 9079P-ME-005; 9079P-NO-005; 9079P-VC-005; 9018P-EE-005; 9018P-EU-005; 9018P-ME-005
Reason for Recall	Safety cap attached to needles within the needle sets may become dislodged exposing the needle and potentially causing the needle to protrude through the packaging. If this issue is not detected, the immediate risk of exposure to the affected devices is needle stick injury to the clinician or healthcare professional. In addition, a puncture of the packaging may compromise the sterility of the needle
Action	1. Kindly check your stock, contact your local agent for remedial action
Product image	
comments	Healthcare professionals are encouraged to report any adverse events Suspected to be associated with the above device or any other medical Device to Director of Medical Device Control contact E-mail dg-padc@moh.gov.om

Directorate General of Pharmaceutical Affairs & Drug Control
Sultanate of Oman



Dr. Mohammed Hamdan Al Rubaie


DIRECTOR GENERAL

3/5/1441 : التاريخ الهجري
29/12/2019 : التاريخ الميلادي
: العرفقات
29-12-19/2823 : رقم

مجلس الصحة
لدول مجلس التعاون
Gulf Health Council



تقارير السلامة للأجهزة والمستلزمات الطبية
Safety Alerts of Medical Device Products

To: HE / Members of the Executive Committee of the Cooperation Council States Members	إلى: سعادة/ أعضاء الهيئة التنفيذية بدول مجلس التعاون.	
Subject:	تقرير سلامة لمنتجات طبية	الموضوع:
Product Name:	Intraosseous vascular access needle sets (Arrow EZ-IO)	اسم المنتج:
Company Name:	Teleflex Medical	اسم الشركة:
Product Photo:		صورة المنتج:
Affected Devices:	9018P-EE-005; 9018P-EU-005; 9018P-ME-005; 9018P-NO-005; 9018P-VC-005; 9001P-EE-005; 9001P-EU-005; 9001P-ME-005; 9001P-NO-005; 9001P-VC-005; 9079P-EE-005; 9079P-EU-005; 9079P-ME-005; 9079P-NO-005; 9079P-VC-005	الأجهزة المتضررة:
Product Reg. Status in GHC:	None	وضع المنتج في مجلس الصحة:
Source of warning:	Teleflex Medical / SFDA	مصدر التحذير:
Reason for warning:	Safety cap attached to needles within the needle sets may become dislodged exposing the needle and potentially causing the needle to protrude through the packaging. If this issue is not detected, the immediate risk of exposure to the affected devices is needle stick injury to the clinician or healthcare professional. In addition, a puncture of the packaging may compromise the sterility of the needle.	سبب التحذير:

3/5/1441 : التاريخ الهجري
29/12/2019 : التاريخ الميلادي
: المرفقات
29-12-19/2823 : رقم

مجلس الصحة
للدول مجلس التعاون
Gulf Health Council



تقارير السلامة للأجهزة والمستلزمات الطبية

Safety Alerts of Medical Device Products

Recommendation:	1- Visually inspect the product. 2- Dispose of product If you identify codes/lots as defective 3- Pass on this notice to all persons who need to be aware within your organization.	التوصيات:
Authorized Representative Details	AR name: First United Medical Company	بيانات التواصل مع الشركة
	Assigned Contact Person: Khalid Mohammad AlFuhaid	
	Mobile/Phone: 0544081009	
	Email: shivappa.gowda@fumedco.net	
Link of Circular	لا يوجد	رابط التعميم:
Any Comment & Report Adverse Reaction:	pms@ghc.sa	للاستفسارات والإبلاغ عن الآثار الجانبية للدواء: