



نتقدم بثقة
Moving Forward
With Confidence



To:

THE DIRECTOR GENERAL OF HEALTH SERVICES IN ALL GOVERNORATES

Commanding Officer, Armed Forces Hospital (Al Khouth & Salalah)

Director General of Engineering Affairs, MOH

Director General of Royal Hospital

Director General of Khoula Hospital

Director General of Medical Supplies (MOH)

Director General of Pvt. Health Est. Affairs (to kindly arrange distribution to all Pvt. Hospitals)

Hospital Director (Al Nahda Hospital)

Hospital Director (Al Massara Hospital)

The Head of Medical Services in SQU Hospital

The Head of Medical Services in Royal Oman Police

The Head of Medical Services in Ministry of Defence

The Head of Medical Services in The Diwan

The Head of Medical Services in The Sultan's Special Force

The Head of Medical Services in Internal Security Services

The Head of Medical Services in Petroleum Development of Oman

The Head of Medical Services in LNG Oman

ALL PRIVATE PHARMACIES & DRUG STORES

After Compliments,

Please find attached our Circular No 72 dated 19/4/2023 Regarding NCMDR Field Safety Corrective Action of Wip'Anios Excel from (mfr: ANIOS Laboratories).

Copy to:

- Director, Office of H.E. The Undersecretary for Health Affairs
- Director of Medical Device Control, DGPA&DC
- Director of Pharmacovigilance & Drug Information Dept, DGPA&DC
- Director of Drug Control Department, DGPA&DC
- Director of Pharmaceutical Licensing Department, DGPA&DC
- Director of Central Quality Control Lab., DGPA&DC
- Supdt. of Central Drug Information



PADC
المديرية العامة للصيدلة والرقابة الدوائية
Directorate General of Pharmaceutical
Affairs & Drug Control



ص.ب: ٣٩٣ مسقط - الرمز البريدي: ١٠٠ - هاتف: ٢٢٣٥٧١١١ - فاكس: ٢٢٣٥٨٤٨٩

P.O. Box: 393 Muscat - Postal Code: 100 - Tel: 22357111 - Fax: 22358489

🐦 dgpa_dc Email: dg-padc@moh.gov.om



Circular No. 72 / 2023

نتقدم بـ
Moving Forward
With Confidence

رؤية عُمان
2040
Oman Vision

28 -09-1444 H

19 -04-2023

Field Safety Corrective Action of Wip'Anios Excel from ANIOS Laboratories.

| | |
|-----------------------|--|
| Source | NCMDR - National Center Medical Device Reporting- SFDA. https://ncmdr.sfda.gov.sa/Secure/CA/CaViewRecall.aspx?caid=4&rid=19512 |
| Product | Wip'Anios Excel. |
| Description | Wipes for Cleaning and disinfection of the medical devices. |
| Manufacturer | ANIOS Laboratories. |
| Local agent | Trade & Marketing Professionals. |
| The affected products | REF: 2446655Y7 (6 x 100 wipes) LOT: E33414S |
| Reason | Presence of Achromobacter sp. |
| Action | 1. Quarantine and to identify and physically isolate the products that you may have at your facility. 2. Contact the local agent for remedial action. |
| comments | Healthcare professionals are encouraged to report any adverse events Suspected to be associated with the above device or any other medical device to Department of Medical Device Control through the E-mail: Med-device@moh.gov.om |



Dr. Mohammed Hamdan Al Rubaie

Director General