



Guideline on Requirements of Class C & D Medical Devices Registration in Sultanate of Oman



DSC
مركز سلامة الدواء
Drug Safety Center

Document Title	Guideline on Requirements of Class C & D Medical Devices Registration in Sultanate of Oman
Document Type	Guideline
Directorate/Institution	Drug Safety Center
Targeted Group	All Local Medical Device Establishment
Document Author	Medical Device Registration Section Task Force
Designation	Medical Device Registration Section Task Force
Document Reviewer	Eng. Faiza Alzadjali
Designation	Director of Medical Device Control Department
Release Date	October 2025
Review Frequency	Every 3 years

Validated by		Approved by	
Name	Ph. Safiya Saleh Al Aghbari	Name	Ph. Ibrahim Nasser Al Rashdi
Designation	Head of Quality Assurance & Safety Management Section	Designation	Directorate General of Drug Safety Center
Signature		Signature	
Date	26/10/2025	Date	18/11/2025

Acknowledgments

The preparation of this guideline was undertaken by the Medical Device Registration Section/ Medical Device Control Department. The Drug Safety Center extends its sincere gratitude to all staff members of the section for their dedicated efforts. Special acknowledgment is given to the members of the Medical Device Registration Section Task Force for their invaluable contributions, in particular:

- Eng. Ghaliya Alqaidi
- Eng. Zahra Alhooti

Table of Contents

Acronyms	4
Definitions	5-6
CHAPTER ONE	
Introduction	7
Purpose	7
Scope	7
Structure	7
CHAPTER TWO	
General Requirements	8-15
CHAPTER THREE	
Responsibilities	16
CHAPTER FOUR	
Document History and Version Control	17
References	17
Annexes	18
Appendix 1	18
Appendix 2	19-37
Appendix 3	38

Acronyms:

MOH	Ministry of Health
DSC	Drug Safety Center
MDCD	Medical Device Control Department
MD	Medical Device
IVD	In Vitro Diagnostic
ISO	International Organization for Standardization
QMS	Quality Management System
GMDN	Global Medical Device Nomenclature
HS code	Harmonized System
AATB certificate	American Association of Tissue Banks Certificate
PSUR	Periodic Safety Update Report
PMCF	Post-market clinical follow-up
CFG	Certificate of foreign government
CDNE	Certificate for medical device not exported from the United States
AIMD	Active Implant Medical Device

Definitions

Medical Device	<p>Medical device' means any instrument, apparatus, implement, machine, appliance, implant, reagent for in vitro use, software, material or other similar or related article, intended by the manufacturer to be used, alone or in combination, for human beings, for one or more of the specific medical purpose(s) of:</p> <ul style="list-style-type: none"> • Diagnosis, prevention, monitoring, treatment or alleviation of disease, • Diagnosis, monitoring, treatment, alleviation of or compensation for an injury, • Investigation, replacement, modification, or support of the anatomy or of a physiological process, • Supporting or sustaining life, • Control of conception, • Disinfection of medical devices, • Providing information by means of in vitro examination of specimens derived from the human body; and does not achieve its primary intended action by pharmacological, immunological or metabolic means, in or on the human body, but which may be assisted in its intended function by such means.
High-Risk Medical Device	High Individual Risk and High Public Health Risk.
Moderate - High Medical devices	Moderate-to high individual Risk and moderate-to high Public Health Risk.
Registration	The process by which a party submits information to the Regulatory Authority in a jurisdiction, regarding the identification and establishment location(s) of the manufacturer and other parties, responsible for supplying a medical device(s) to the market in that jurisdiction.

Label	Means the written, printed or graphic information appearing either on the device itself, or on the packaging of each unit or on the packaging of multiple devices.
Manufacturer	Means any natural or legal person with responsibility for design and/or manufacture of a medical device with the intention of making the medical device available for use, under his name; whether or not such a medical device is designed and/or manufactured by that person himself or on his behalf by another person(s).
Accessories	Means a product intended specifically by its manufacturer to be used together with a medical device to enable that medical device to achieve its intended purpose.
Accessories to a medical device	Means an article intended specifically by its manufacturer to be used together with a particular medical device to enable or assist that device to be used in accordance with its intended use
Verification	Confirmation by examination and provision of objective evidence that the specified requirements have been fulfilled.
Validation	Validation means confirmation by examination and provision of objective evidence that the particular requirements for a specific intended use can be consistently fulfilled.
Risk	Combination of the probability of occurrence of harm and the severity of that harm.
Intended use / purpose	The objective intent of the manufacturer regarding the use of a product, process or service as reflected in the specifications, instructions and information provided by the manufacturer.

CHAPTER ONE

Introduction

Medical devices are required to be registered to cater the regulatory requirements. Regulation is enforced in Oman via ministerial decree 113/2020. This will ensure that products entering the market are safe and efficient.

This guidance document is meant to assist applicants for the registration of moderate –high & high risk medical devices in sultanate of Oman.

Applicants are strongly encouraged to familiarize themselves with the criteria and requirements for review processes outlined in this guidance and the other relevant guidance documents before submitting their applications.

Purpose

The purpose of this guidance is to describe the procedures and general requirements for the submission of moderate –high- & high-risk medical device registration.

Scope

This guidance applies to the following products:

- High Risk Medical Devices. (Class D) -***All other class III/ Class IV/ AIMD***
- Moderate - High Risk Medical Devices. (Class C)- ***Class IIb/ Class III***

Structure

This is the first version of this guidance and it consists of several chapters. Chapter one covers a brief introduction to the guideline as well as the purpose, scope and structure. Chapter two explains the general requirements of medical device registration. Chapter three covers the requirements of moderate –high & high risk medical device registration. Chapter four comprises of the document history and version control table, references and appendixes.

CHAPTER TWO

General Requirements

1. Registration requirements shall be presented in a clear, organized, readily, searchable and unambiguous manner.
2. Apply for registration through the online portal.
<https://moh.gov.om/en/services/businesses/drug-safety-center/request-for-registration-of-medical-device/>
3. To fulfil the prerequisite of the MD registration, applicant must complete, the Initial approval of wholesale Activity, local establishment approval and Manufacturer registration.
4. Payment of fees should be made according to Risk-classification.
5. Applications with the incorrect risk classification of devices may result in the re- submission of the applications according to the appropriate risk class.

Note:

- **Drug Safety Center has the right to request more requirements as per product type if needed.**
- **In case of any variation in the Medical Devices the manufacturer shall apply through variation services after the device registration is completed.**
- **Medical Device Dossier File Sections Requirements:**
 - High risk medical device (all sections required).
 - Moderate - High risk device (all sections except 5 & 6).

Requirements of moderate - high- & high-risk medical device registration:

Section 1: Application type

This section includes the type of application whether the device is Medical device or IVD, name of the device, the risk classification of the device, device grouping, device Category and which jurisdiction the device follow.

Regulatory jurisdiction the device follows:

- **KSA (MDMA)**
- **USA**
- **EU**
- **Canada**
- **Australia**
- **Japan**
- **Others (specify)**

Section 2: Manufacturer information

This section includes name, physical site address and details of legal manufacturer, quality management system used and the certificates e.g. (ISO 13485) for the manufacturer. And the following:

- Legal manufacturer Details (Legal manufacturer Name, Manufacturer registration status, Legal manufacturer country/city, Tel no, Fax no, Email, QMS Certificate Number, and QMS Certificate expiry date)
- All information will be gathered from previous step which is MD manufacturer registration service.
- **For more details, refer to the guideline GD 22** (Guidance Document GD 22: Requirements of Medical Device Manufacturer Registration in Sultanate of Oman).

Section 3: Medical Device information

This section defines the medical device and accessories information, and medical device grouping/bundling. The following points shall be submitted.

Medical Device information:

- Trade / Brand name.
- Model name / number.
- Serial number.
- Medical device risk classification.
- Intended use.
- Medical device category.
- Manufacturer device ID (identification number).
- Is Single Use Device?
- Is this Product Sterile?
- Countries Where the Device Used.
- GMDN.
- Nomenclature code if different than GMDN.
- HS code.
- Method of Traceability.
- Shelf life in months (if applicable)
- Storage condition. (Manufacturer recommendation).
- First year sold.
- Warnings.
- Active ingredient.
- Principles of operation/mode of action (how it works/ operates).
- Picture or drawing of the device which should be details (include sufficient explanation to understand the drawing).
- Description of any devices required to operate the device (IT infrastructure, laptop, mobile smart phone) if applicable.
- Declaration (does the device require any specific declaration to be made such as containing animal tissue / medicines/human blood derivatives)

Medical device grouping/ bundling:

This field describes which grouping / bundling to be filled as per Annex (1), 50 product maximum in one application, and it include Product name, manufacturer name, trade brand name, model number, risk classification and category. Refer to bundling and grouping guidance for more information.

- Medical Devices Single
- Medical Devices Family
- Medical Devices System
- Medical Devices Procedure Pack

Note: If the device has accessories, they may be included with the device within the same registration application, unless they are marketed separately

Medical Device Accessories information:

In case there are any accessories it will including the following:

- Accessories Name.
- Risk classification.
- Serial Number.
- Description of accessories.
- GMDN.
- Model No.
- Device Identifier.
- Batch/ Lot/ reference no.
- Label.
- Catalogue.
- IFU.

Section 4: Device Labeling

This section of registration requirements shall include a full set of:

- Instruction for Use (IFU).
- Label on the Device.
- Product Labels and Packaging
- Promotional materials /catalogue

Section 5: Essential Principles & evidence of conformity

This section intended to demonstrate how the manufacturer has met the essential principles of safety and performance. The Essential Principal Checklists Annex (2). Can be used to demonstrate this and should form this section of the registration requirements. Alternative formats are allowed as long as all essential principles have been considered as follows:

- All applicable essential principles have been identified and conformity has been validated by evidenced.
- All non-applicable essential principles have a rationale detailing why they are not applicable.
- The precise identity of the controlled documents offering evidence of conformity with each essential principle is documented.

Section 6: Summary of design Verification & Validation

This section provides details on the device design and how it is manufactured. This section shall include:

- **Design Stages**
 - Reference to design procedure.
 - Description of design stages.
 - Confirmation of the verification and validation conducted on the device .

- **Manufacturing processes**
 - Manufacturing information to include:
 - Manufacturing process flow
 - Manufacturing specifications including in process testing
 - Final inspection and acceptance criteria
 - Manufacturing validation
- **Manufacturing structure**
 - Name and address of place where design was carried out (all sites and subcontractors)
 - Name and address of place where manufacturing is carried out (all sites and subcontractors)
 - Detail of suppliers including identification of critical suppliers
 - Confirmation of subcontractor contracts (if applicable)

Section 7: Product Verification and Validation

This section provides product certificates, if product follow EU jurisdiction:

- **Moderate-High risk (Class C)**

Class IIb/ Class III:

- CE /EC Certificate.
- Full quality assurance / Equivalent.
- Design examination certificate / Equivalent.
- Others.

- **High risk (Class D)**

All other class III/ Class IV/ AIMD

- CE /EC Certificate.
- Full quality assurance / Equivalent.
- Design examination certificate / Equivalent.
- TSE free certificate (if biological).
- Others.

This section provides product certificates if product follow US jurisdiction:

- **Moderate-High Risk (Class C)**

Class IIb/ Class III:

- 510k .
- PMA.
- Others.

- **High Risk (Class D)**

All other class III/ Class IV/ AIMD

- PMA.
- AATB certificate (if from US) (if biological).
- Others.

Section8: Clinical evidence

This section provides clinical evidence about the products and shall include:

- Clinical evaluation report.
- Stability testing.
- Shelf life testing.
- Software validation (If applicable).

Section 9: Post Market control

Medical device manufacturers shall prepare a periodic safety update report (PSUR) for each device and where relevant for each category or group of devices summarizing the results and conclusions of the analyses of the post-market surveillance data gathered as a result of the post-market surveillance plan together with a rationale and description of any preventive and corrective actions taken. Post-market Surveillance Plan is a proactive and systematics process to collect information.

Throughout the lifetime of the device concerned, that PSUR shall set out:

- The conclusions of the benefit-risk determination
- The main findings of the Post-market clinical follow-up (PMCF)
- The volume of sales of the device and an estimate of the size and other characteristics of the population using the device and, where practicable, the usage frequency of the device.
- Manufacturers of high risk devices (C&D) shall update the PSUR at least annually.

Section 10: Status of device distribution

In this section the establishment of medical devices should:

- List the countries where the device have been marketed with evidence and /or a letter issued by the legal manufacturer conforming the product distribution list.
- Submit free sale certificate from competent authority in origin country or certificate of foreign government CFG if the product from USA and CDNE (certificate for medical device not exported from the United States).

Section 11: Declaration of conformity

This section of registration requirements includes declaration of conformity from manufacturer and it is containing:

- Product Name
- Model number
- Classification
- Statement that the declaration is issued under the sole responsibility of the manufacturer
- Issued/signed stamped from manufacturer.

Section 12 : Benefit-Risk Analysis and risk management

Note: All Dossier sections requirements should be documented and filled in the online platform.

CHAPTER THREE

Responsibilities:

Department and section staff	Support the review, evaluation, and processing of Moderate-High and High-Risk Medical Device registration applications in accordance with the applicable regulatory requirements.
Department Directors and Section Heads	Supervise and ensure compliance with the registration procedures, approve recommendations, and provide oversight for all Moderate-High and High-Risk Medical Device registration activities.
QASM Section	Monitor quality assurance aspects, safety, and post-market surveillance related to registered Moderate-High and High-Risk Medical Devices.
DG-DSC	Oversee implementation of the regulatory framework, ensure alignment with national and international standards, and endorse final decisions on Moderate-High and High-Risk Medical Device registrations.

CHAPTER FOUR

Document History and Version Control

Version	Description	Review Date
1	Initial Release	October 2025
2		
3		

References:

Asian Harmonization Working Party (AHWP), 2015. *Guidance for Preparation of a Common Submission Dossier Template (CSDT) Dossier for General Medical Device Product Submission (AHWP/WG1/F001:2015)*.

International Medical Device Regulators Forum (IMDRF), 2024. *Essential Principles of Safety and Performance of Medical Devices and IVD Medical Devices (IMDRF/GRRP WG/N47 FINAL:2024)*.

World Health Organization (WHO), n.d. *Medical Device Definition*. Geneva: World Health Organization.

Global Harmonization Task Force (GHTF), 2010. *Registration of Manufacturers and Listing of Medical Devices (GHTF/SG1/N065:2010)*.

International Medical Device Regulators Forum (IMDRF), 2024. *Principles of Labeling for Medical Devices and IVD Medical Devices (IMDRF/GRRP WG/N52 FINAL:2024)*.

Global Harmonization Task Force (GHTF), 2012. *Definitions of the Terms Manufacturer, Authorized Representative, Distributor and Importer (GHTF/SG1/N071:2012)*.

Global Harmonization Task Force (GHTF), 2012. *Principles of Medical Devices Classification (GHTF/SG1/N77:2012)*.

Annexes

Appendix 1: Medical Device Grouping /Bundling

Please refer to the Guidance Document on grouping & bundling for medical devices

	A	B	C	D	E	F	G
2							
3				Medical Device Grouping\ Bundling:			
4				(choose from drop down list)			
5							
6	Sr. No	Product Description	Intended Purpose	Category	Classification	Trade/Brand Name	Model Number
7							
8							
9							
10							
11							
12							
13							
14							
15							
16							
17							
18							
19							
20							
21							
22							
23							

Appendix 2: Essential Principles of Safety and Performance for Medical Devices

Essential Principal Checklist	
Device:	

Essential Principal	Applicable to the device?	Standards Applied	Method of Conformity	Identity of Specific Documents
5. General Requirements				
5.1 Medical devices should be designed and manufactured in such a way that, when used under the conditions and for the purposes intended and, where applicable, by virtue of the technical knowledge, experience, education or training of intended users, they will not compromise the clinical condition or the safety of patients, or the safety and health of users or, where applicable, other persons, provided that any risks which may be associated with their use constitute acceptable risks when weighed against the benefits to the patient and are compatible with a high level of protection of health and safety.				

Essential Principal	Applicable to the device?	Standards Applied	Method of Conformity	Identity of Specific Documents
<p>5.2 The solutions adopted by the manufacturer for the design and manufacture of the devices should conform to safety principles, taking account of the generally acknowledged state of the art. When risk reduction is required, the manufacturer should control the risk(s) so that the residual risk(s) associated with each hazard is judged acceptable. The manufacturer should apply the following principles in the priority order listed:</p> <ul style="list-style-type: none"> ▪ identify known or foreseeable hazards and estimate the associated risks arising from the intended use and foreseeable misuse, ▪ eliminate risks as far as reasonably practicable through inherently safe design and manufacture, ▪ reduce as far as is reasonably practicable the remaining risks by taking adequate protection measures, including alarms, ▪ Inform users of any residual risks. 				
<p>5.3 Devices should achieve the performance intended by the manufacturer and be designed, manufactured and packaged in such a way that they are suitable for one or more of the functions within the scope of the definition of a medical device applicable in each jurisdiction.</p>				

Essential Principal	Applicable to the device?	Standards Applied	Method of Conformity	Identity of Specific Documents
5.4 The characteristics and performances referred to in Clauses 5.1, 5.2 and 5.3 should not be adversely affected to such a degree that the health or safety of the patient or the user and, where applicable, of other persons are compromised during the lifetime of the device, as indicated by the manufacturer, when the device is subjected to the stresses which can occur during normal conditions of use and has been properly maintained in accordance with the manufacturer's instructions.				
5.5 The devices should be designed, manufactured and packed in such a way that their characteristics and performances during their intended use will not be adversely affected under transport and storage conditions (for example, fluctuations of temperature and humidity) taking account of the instructions and information provided by the manufacturer.				
5.6 The benefits must be determined to outweigh any undesirable side effects for the performances intended.				
Design and Manufacturing Requirements				
5.7 Chemical, physical and biological properties				

Essential Principal	Applicable to the device?	Standards Applied	Method of Conformity	Identity of Specific Documents
<p>The devices should be designed and manufactured in such a way as to ensure the characteristics and performance referred to in Clauses 5.1 to 5.6 of the 'General Requirements'. Particular attention should be paid to:</p> <ul style="list-style-type: none"> the choice of materials used, particularly as regards toxicity and, where appropriate, flammability, the compatibility between the materials used and biological tissues, cells, body fluids, and specimens, taking account of the intended purpose of the device, The choice of materials used should reflect, where appropriate, matters such as hardness, wear and fatigue strength. 				
<p>The devices should be designed, manufactured and packed in such a way as to minimize the risk posed by contaminants and residues to the persons involved in the transport, storage and use of the devices and to patients, taking account of the intended purpose of the product. Particular attention should be paid to tissues exposed and to the duration and frequency of exposure.</p>				

Essential Principal	Applicable to the device?	Standards Applied	Method of Conformity	Identity of Specific Documents
The devices should be designed and manufactured in such a way that they can be used safely with the materials, substances and gases with which they enter into contact during their normal use or during routine procedures; if the devices are intended to administer medicinal products they should be designed and manufactured in such a way as to be compatible with the medicinal products concerned according to the provisions and restrictions governing these products and that their performance is maintained in accordance with the intended use.				
Where a device incorporates, as an integral part, a substance which, if used separately, may be considered to be a medicinal product/drug as defined in the relevant legislation that applies within that jurisdiction and which is liable to act upon the body with action ancillary to that of the device, the safety, quality and usefulness of the substance should be verified, taking account of the intended purpose of the device.				
The devices should be designed and manufactured in such a way as to reduce as far as reasonably practicable and appropriate the risks posed by substances that may leach or leak from the device.				

Essential Principal	Applicable to the device?	Standards Applied	Method of Conformity	Identity of Specific Documents
Devices should be designed and manufactured in such a way as to reduce as far as reasonably practicable and appropriate risks posed by the unintentional ingress or egress of substances into or from the device taking into account the device and the nature of the environment in which it is intended to be used.				
5.8 Infection and microbial contamination				
<p>The devices and manufacturing processes should be designed in such a way as to eliminate or to reduce as far as reasonably practicable and appropriate the risk of infection to patients, users and, where applicable, other persons. The design should:</p> <ul style="list-style-type: none"> allow easy handling, <p>and, where necessary:</p> <ul style="list-style-type: none"> reduce as far as reasonably practicable and appropriate any microbial leakage from the device and/or microbial exposure during use, Prevent microbial contamination of the device, or specimen where applicable, by the patient, user or other person. 				

Essential Principal	Applicable to the device?	Standards Applied	Method of Conformity	Identity of Specific Documents
Where a device incorporates substances of biological origin, the risk of infection must be reduced as far as reasonably practicable and appropriate by selecting appropriate sources, donors and substances and by using, as appropriate, validated inactivation, conservation, and test and control procedures.				
In some jurisdictions products incorporating tissues, cells and substances of non-human origin may be considered medical devices. In this case, such tissues, cells and substances should originate from animals that have been subjected to veterinary controls and surveillance adapted to the intended use of the tissues. National regulations may require that the manufacturer and/or the Regulatory Authority retain information on the geographical origin of the animals. Processing, preservation, testing and handling of tissues, cells and substances of animal origin should be carried out so as to provide optimal safety. In particular, safety with regard to viruses and other transmissible agents should be addressed by implementation of validated methods of elimination or inactivation in the course of the manufacturing process.				

Essential Principal	Applicable to the device?	Standards Applied	Method of Conformity	Identity of Specific Documents
In some jurisdictions products incorporating human tissues, cells and substances may be considered medical devices. In this case, the selection of sources, donors and/or substances of human origin, the processing, preservation, testing and handling of tissues, cells and substances of such origin should be carried out so as to provide optimal safety. In particular, safety with regard to viruses and other transmissible agents should be addressed by implementation of validated methods of elimination or inactivation in the course of the manufacturing process.				
Devices labelled as having a special microbiological state should be designed, manufactured and packed to ensure they remain so when placed on the market and remain so under the transport and storage conditions specified by the manufacturer.				
Devices delivered in a sterile state should be designed, manufactured and packed in a non-reusable pack, and/or according to appropriate procedures, to ensure that they are sterile when placed on the market and remain sterile, under the transport and storage conditions indicated by the manufacturer, until the protective packaging is damaged or opened.				
Devices labelled either as sterile or as having a special microbiological state should have been processed, manufactured and, if applicable, sterilized by appropriate, validated methods.				
Devices intended to be sterilized should be manufactured in appropriately controlled (e.g. environmental) conditions.				

Essential Principal	Applicable to the device?	Standards Applied	Method of Conformity	Identity of Specific Documents
Packaging systems for non-sterile devices should keep the product without deterioration at the level of cleanliness stipulated and, if the devices are to be sterilized prior to use, minimize the risk of microbial contamination; the packaging system should be suitable taking account of the method of sterilization indicated by the manufacturer.				
The packaging and/or label of the device should distinguish between identical or similar products placed on the market in both sterile and non-sterile condition.				
5.9 Manufacturing and environmental properties				
If the device is intended for use in combination with other devices or equipment, the whole combination, including the connection system should be safe and should not impair the specified performance of the devices. Any restrictions on use applying to such combinations should be indicated on the label and/or in the instructions for use.				

Essential Principal	Applicable to the device?	Standards Applied	Method of Conformity	Identity of Specific Documents
<p>Devices should be designed and manufactured in such a way as to remove or reduce as far as reasonably practicable and appropriate:</p> <ul style="list-style-type: none"> the risk of injury, in connection with their physical features, including the volume/pressure ratio, dimensional and where appropriate ergonomic features; risks connected with reasonably foreseeable external influences or environmental conditions, such as magnetic fields, external electrical and electromagnetic effects, electrostatic discharge, pressure, humidity, temperature or variations in pressure and acceleration; the risks connected to their use in conjunction with materials, substances and gases with which they may come into contact during normal conditions of use; the risks of accidental penetration of substances into the device; the risk of incorrect identification of specimens; the risks of reciprocal interference with other devices normally used in the investigations or for the treatment given; Risks arising where maintenance or calibration are not possible (as with implants), from ageing of materials used or loss of accuracy of any measuring or control mechanism. 				

Essential Principal	Applicable to the device?	Standards Applied	Method of Conformity	Identity of Specific Documents
Devices should be designed and manufactured in such a way as to minimize the risks of fire or explosion during normal use and in single fault condition. Particular attention should be paid to devices whose intended use includes exposure to or use in association with flammable substances or substances which could cause combustion.				
Devices must be designed and manufactured in such a way as to facilitate the safe disposal of any waste substances.				
5.10 Devices with a diagnostic or measuring function				
Devices with a measuring function, where inaccuracy could have a significant adverse effect on the patient, should be designed and manufactured in such a way as to provide sufficient accuracy, precision and stability for their intended purpose of the device. The limits of accuracy should be indicated by the manufacturer.				
Diagnostic devices should be designed and manufactured in such a way as to provide sufficient accuracy, precision and stability for their intended use, based on appropriate scientific and technical methods. In particular the design should address sensitivity, specificity, trueness, repeatability, and reproducibility, control of known relevant interference and limits of detection, as appropriate.				

Essential Principal	Applicable to the device?	Standards Applied	Method of Conformity	Identity of Specific Documents
Where the performance of devices depends on the use of calibrators and/or control materials, the traceability of values assigned to such calibrators and/or control materials should be assured through a quality management system.				
Any measurement, monitoring or display scale should be designed in line with ergonomic principles, taking account of the intended purpose of the device.				
Wherever possible values expressed numerically should be in commonly accepted, standardised units, and understood by the users of the device. Note: While SG1 generally supports convergence on the global use of internationally standardised measurement units, considerations of safety, user familiarity, and established clinical practice may justify the use of other recognised measurement units.				
5.11 Protection against radiation				
General				
5.11.1.1 Devices should be designed and manufactured and packaged in such a way that exposure of patients, users and other persons to any emitted radiation should be reduced as far as practicable and appropriate, compatible with the intended purpose, whilst not restricting the application of appropriate specified levels for therapeutic and diagnostic purposes.				
Intended radiation				

Essential Principal	Applicable to the device?	Standards Applied	Method of Conformity	Identity of Specific Documents
5.11.2.1 Where devices are designed to emit hazardous, or potentially hazardous, levels of visible and/or invisible radiation necessary for a specific medical purpose the benefit of which is considered to outweigh the risks inherent in the emission, it should be possible for the user to control the emissions. Such devices should be designed and manufactured to ensure reproducibility of relevant variable parameters within an acceptable tolerance.				
5.11.2.2 Where devices are intended to emit potentially hazardous, visible and/or invisible radiation, they should be fitted, where practicable, with visual displays and/or audible warnings of such emissions.				
Unintended radiation				
5.11.3.1 Devices should be designed and manufactured in such a way that exposure of patients, users and other persons to the emission of unintended, stray or scattered radiation is reduced as far as practicable and appropriate.				
Instructions for use				
5.11.4.1 The operating instructions for devices emitting radiation should give detailed information as to the nature of the emitted radiation, means of protecting the patient and the user and on ways of avoiding misuse and of eliminating the risks inherent in installation.				

Essential Principal	Applicable to the device?	Standards Applied	Method of Conformity	Identity of Specific Documents
Ionizing radiation				
5.11.5.1 Devices intended to emit ionizing radiation should be designed and manufactured in such a way as to ensure that, where practicable, the quantity, geometry and energy distribution (or quality) of radiation emitted can be varied and controlled taking into account the intended use.				
5.11.5.2 Devices emitting ionizing radiation intended for diagnostic radiology should be designed and manufactured in such a way as to achieve appropriate image and/or output quality for the intended medical purpose whilst minimising radiation exposure of the patient and user.				
5.11.5.3 Devices emitting ionizing radiation, intended for therapeutic radiology should be designed and manufactured in such a way as to enable reliable monitoring and control of the delivered dose, the beam type and energy and where appropriate the energy distribution of the radiation beam.				
5.12 Requirements for medical devices connected to or equipped with an energy source				

Essential Principal	Applicable to the device?	Standards Applied	Method of Conformity	Identity of Specific Documents
5.12.1 Devices incorporating electronic programmable systems, including software, should be designed to ensure the repeatability, reliability and performance of these systems according to the intended use. In the event of a single fault condition in the system, appropriate means should be adopted to eliminate or reduce as far as practicable and appropriate consequent risks.				
Devices where the safety of the patients depends on an internal power supply should be equipped with a means of determining the state of the power supply.				
Devices where the safety of the patients depends on an external power supply should include an alarm system to signal any power failure.				
Devices intended to monitor one or more clinical parameters of a patient should be equipped with appropriate alarm systems to alert the user of situations which could lead to death or severe deterioration of the patient's state of health				
Devices should be designed and manufactured in such a way as to reduce as far as practicable and appropriate the risks of creating electromagnetic interference which could impair the operation of this or other devices or equipment in the usual environment.				
Devices should be designed and manufactured in such a way as to provide an adequate level of intrinsic immunity to electromagnetic disturbance to enable them to operate as intended.				

Essential Principal	Applicable to the device?	Standards Applied	Method of Conformity	Identity of Specific Documents
<p>Protection against electrical risks</p> <p>Devices should be designed and manufactured in such a way as to avoid, as far as possible, the risk of accidental electric shocks during normal use and in single fault condition, provided the devices are installed and maintained as indicated by the manufacturer.</p>				
<p>5.13 Protection against mechanical risks</p>				
<p>Devices should be designed and manufactured in such a way as to protect the patient and user against mechanical risks connected with, for example, resistance to movement, instability and moving parts.</p>				
<p>Devices should be designed and manufactured in such a way as to reduce to the lowest practicable level the risks arising from vibration generated by the devices, taking account of technical progress and of the means available for limiting vibrations, particularly at source, unless the vibrations are part of the specified performance.</p>				
<p>Devices should be designed and manufactured in such a way as to reduce to the lowest practicable level the risks arising from the noise emitted, taking account of technical progress and of the means available to reduce noise, particularly at source, unless the noise emitted is part of the specified performance.</p>				

Essential Principal	Applicable to the device?	Standards Applied	Method of Conformity	Identity of Specific Documents
Terminals and connectors to the electricity, gas or hydraulic and pneumatic energy supplies which the user has to handle should be designed and constructed in such a way as to minimize all possible risks.				
Accessible parts of the devices (excluding the parts or areas intended to supply heat or reach given temperatures) and their surroundings should not attain potentially dangerous temperatures under normal use.				
5.14 Protection against the risks posed to the patient by supplied energy or substances				
Devices for supplying the patient with energy or substances should be designed and constructed in such a way that the delivered amount can be set and maintained accurately enough to guarantee the safety of the patient and of the user.				
Devices should be fitted with the means of preventing and/or indicating any inadequacies in the delivered amount which could pose a danger. Devices should incorporate suitable means to prevent, as far as possible, the accidental release of dangerous levels of energy from an energy and/or substance source.				
The function of the controls and indicators should be clearly specified on the devices. Where a device bears instructions required for its operation or indicates operating or adjustment parameters by means of a visual system, such information should be understandable to the user and, as appropriate, the patient.				

Essential Principal	Applicable to the device?	Standards Applied	Method of Conformity	Identity of Specific Documents
5.15 Protection against the risks posed to the patient for devices for self-testing or self-administration				
Such devices should be designed and manufactured in such a way that they perform appropriately for their intended purpose taking into account the skills and the means available to users and the influence resulting from variation that can reasonably be anticipated in user's technique and environment. The information and instructions provided by the manufacturer should be easy for the user to understand and apply.				
Such devices should be designed and manufactured in such a way as to reduce as far as practicable the risk of use error in the handling of the device and, if applicable, the specimen, and also in the interpretation of results.				
Such devices should, where reasonably possible, include a procedure by which the user can verify that, at the time of use that the product will perform as intended by the manufacturer.				
5.16 Information supplied by the manufacturer				

Essential Principal	Applicable to the device?	Standards Applied	Method of Conformity	Identity of Specific Documents
<p>5.16.1 Users should be provided with the information needed to identify the manufacturer, to use the device safely and to ensure the intended performance, taking account of their training and knowledge. This information should be easily understood.</p> <p>Note: Further information is provided in <i>SG1/N009 Labelling for Medical Devices</i> and in <i>SG1/N043 Labelling for Medical Devices (revised)</i>.</p>				
<p>5.17 Performance evaluation including, where appropriate, clinical evaluation</p>				
<p>All data generated in support of performance evaluation should be obtained in accordance with the relevant requirements applicable in each jurisdiction.</p>				
<p>Clinical investigations on human subjects should be carried out in accordance with the spirit of the Helsinki Declaration. This includes every step in the clinical investigation from first consideration of the need and justification of the study to publication of the results. In addition, some countries may have specific regulatory requirements for pre-study protocol review or informed consent.</p>				

Appendix 3: Procedure Flowchart

