



To:

THE DIRECTOR GENERAL OF HEALTH SERVICES IN ALL GOVERNORATES
Commanding Officer, Armed Forces Hospital (Al Khoudh & Salalah)
Director General of Engineering Affairs, MOH
Director General of Royal Hospital
Director General of Khoula Hospital
Director General of Medical Supplies (MOH)
Director General of Pvt. Health Est. Affairs (to kindly arrange distribution to all Pvt. Hospitals)
Hospital Director (Al Nahda Hospital)
Hospital Director (Al Massara Hospital)
The Head of Medical Services in SQU Hospital
The Head of Medical Services in Royal Oman Police
The Head of Medical Services in Ministry of Defence
The Head of Medical Services in The Diwan
The Head of Medical Services in The Sultan's Special Force
The Head of Medical Services in Internal Security Services
The Head of Medical Services in Petroleum Development of Oman
The Head of Medical Services in LNG Oman
ALL PRIVATE PHARMACIES & DRUG STORES

After Compliments,

Please find attached our Circular No 226 dated 23/10/2023 Regarding NCMDR recall of Radox Urine Liquid Control Level 2 from (mfr: Radox Laboratories Ltd).

Copy to:

- Director, Office of H.E. The Undersecretary for Health Affairs
- Director of Medical Device Control, DGPA&DC
- Director of Pharmacovigilance & Drug Information Dept, DGPA&DC
- Director of Drug Control Department, DGPA&DC
- Director of Pharmaceutical Licensing Department, DGPA&DC
- Director of Central Quality Control Lab., DGPA&DC
- Supdt. of Central Drug Information





Circular No. 226/2023

نتقدم بثقة
Moving Forward
With Confidence

رؤية عمان
2040
Oman Vision

08 -04-1445 H

23 -10-2023

Recall of Randox Urine Liquid Control Level 2.from Randox Laboratories Ltd.

Source	NCMDR- National Center for Medical Devices Reporting- SFDA https://ncmdr.sfda.gov.sa/Secure/CA/CaViewRecall.aspx?caid=4&rid=19714
Product	Randox Urine Liquid Control Level 2.
Description	IVD; In vitro diagnostic devices.
Manufacturer	Randox Laboratories Ltd .
Local Agent	Mustafa Sultan Science & Industry Co.LLC.
The affected products	Catalog Number: UC5074 Code Information GTIN: 05055273207569 Batch/Lot Number: 1209UC Expiry Date: 28 Mar 24 Manufacturing Date: 28 Apr 22
Reason	There is vial to vial variation resulting in some vials of the above product recovering positive for hCG. There has been a transcription error for Creatinine in the Instructions For Use (IFU) of the above product. The target and ranges for Creatinine for the Roche Creatinine Plus method have been listed incorrectly.
Action	1. Discontinue use of and discard any of the above immediately. 2. Contact the local agent for remedial action.
comments	Healthcare professionals are encouraged to report any adverse events Suspected to be associated with the above device or any other medical device to Department of Medical Device Control through the E-mail: Med-device@moh.gov.om

Dr. Mohammed Hamdan Al Rubaie

Director General

